

Certificate Of Calibration

for
ABC Company

Cust. P.O. #: 12345678
Test Item: TJ36-NNXL-316U-18-SMPW-
M-CAL-3

Report #: OM- 121125534
WO : SC001258769

Recal Date: Per System Application

Ref. I.D.: HVN140319-01

CAL-3

Omega Engineering, Inc. certifies that the above instrumentation has been calibrated and tested to **meet or exceed** the published specifications. This calibration and testing was performed using instrumentation and standards that are traceable to the **National Institute of Standards and Technology**. Calibration has been performed in compliance with ISO 10012-1, ISO 9001 and ANSI/NCSL Z540-1-1994 as well as ASTM E 230 and ANSI MC96.1. This Certificate/Report shall not be reproduced, except in full, without written consent of Omega Engineering Inc.

Test Conditions: Temperature 70°F Relative Humidity 52%

Procedure used: QAP-2100

The maximum calibration uncertainty is calculated to be 0.3C from -25C to 500 C and 0.55C from 500C to 1100C.

INSTRUMENTS USED:

<u>MODEL</u>	<u>SERIAL#</u>	<u>CAL DUE DATE</u>	<u>N.I.S.T. NUMBERS</u>
Agilent 34401A	MY41015106	04/10/20	NNDM-100-11
TRCIII	I-0014	03/14/20	NNIP-100-07
DP251	1360-007-924	07/25/19	NNDP-100-12
DP251	2193-022-2473	08/24/19	NNDP-100-10
RTD(Burns)	745098	08/24/19	NNPR-100-01
RTD(Burns)	765565	08/24/19	NNPR-100-05
RTD (Burns)	765566	08/24/19	NNPR-100-06

Probe No.	Nominal Temperature	Actual Test Temperature	Indicated Temperature
1	212 °F	212.0 °F	211.9 °F
1	392 °F	392.0 °F	391.1 °F
1	752 °F	752.0 °F	749.9 °F

Metrology Technician
Calibration Date: 06/19/19

Quality Assurance Inspector
Page 1 of 1