





HH20CAL

Economical Thermocouple Calibrator



M1041/0497



UNPACKING

Remove the Packing List and verify that you have received all equipment, including the following (quantities in parentheses): calibrator (1), 9V battery (1), and Instruction Sheet (1). If you have any questions about the shipment, please call the OMEGA Customer Service Department.

When you receive the shipment, inspect the container and equipment for signs of damage. Note any evidence of rough handling in transit. Immediately report any damage to the shipping agent.

NOTE

The carrier will not honor damage claims unless all shipping material is saved for examination. After examining and removing contents, save packing material and carton in the event reshipment is necessary.

GENERAL DESCRIPTION

The OMEGA® HH20CAL Economical Thermocouple Calibrator is used with OMEGA HH20 Series; HH25, 26; HH55; HH81, 82; HH95, 99; 450 Series, 871, 872 Handheld Thermometers. The HH20CAL is designed to plug directly onto the OMEGA Handhelds, enabling them to be used as portable calibrators for digital readouts, controllers and other instrumentation.

The calibrator will simulate J, K, or T thermocouples, as selected by the thermocouple selector switch. The simulated temperature can be adjusted anywhere between -350°F and 2500°F using the coarse and fine tune dials and the HIGH/MEDIUM/LOW range switch. The HH20CAL allows quick, easy adjustment to within 0.1°, overcoming sensitivity problems encountered with other manufacturer's models. The output is compensated to correct for ambient temperature, ensuring repeatable calibration points even in drafts and other temperature changes.

BATTERY REPLACEMENT

The HH20CAL operates on a 9 volt alkaline battery providing approximately 1000 hours of operation and easy replacement (Open case by removing the three screws on the back of the unit). The battery condition is checked when the calibrator is off. A low battery is indicated by a reference thermometer reading of less than 100°C (212°F). A new battery will display approximately 230°C.

OPERATION

1. Plug the HH20CAL onto the OMEGA Series Handheld Thermometer being used.
2. Connect the proper thermocouple wire and miniature male SMP connector to the HH20CAL.
3. Connect the other end of the thermocouple wire to the instrument to be calibrated.
4. Switch on the OMEGA Series Handheld and the HH20CAL by moving the slide switch to J, K or T.
5. Using the HIGH/MEDIUM/LOW range switch and the COARSE/FINE adjust knobs set the HH20CAL to the output required as read on the Handheld.

NOTE

In order to make the HH20CAL completely compatible with a wide range of thermometers, the male connector on the bottom of the unit is fitted with a universal mounting arrangement. The center position mates with single input thermometers while the offset position mates with dual input models. To change connector positions, disconnect the unit from any instrument and remove the back support thumbscrew. Open the case by removing the three screws on the back of the unit. After removing the case back, the connector is accessible. Simply lift the connector and its mounting pin from the stud and press the pin into the other stud. The connector is then press-fitted into place on the pin. Reassemble the case and back support.

SPECIFICATIONS

Accuracy: $\pm(0.3\% + 1^{\circ}\text{C})$ + accuracy of indicating thermometer, (includes ambient temperature effect).

Range: -350°F to 2500°F overall (-212°C to 1371°C)

Low Range: -350 to 250°F

Medium Range: 200 to 1200°F

High Range: 1200 to 2500°F

Resolution: Allows tuning to 0.1°C (with a 0.1°C indicating thermometer)

Operating Ambient: 32°F to 130°F less than 80% relative humidity up to 95°F

Storage Temperature: -30°F to 140°F less than 80% R.H. up to 95°F

Connections: Subminiature OMEGA SMP thermocouple plug and jack.

Power: Standard 9 volt battery

Dimensions (L x W x H): 127 x 71 x 38 mm (5" x 2.8" x 1.5")

Weight: 227 g (8 oz)



Ω OMEGA®

OMEGAnet™ On-Line Service
http://www.omega.com

Internet e-mail
info@omega.com

Servicing North America:

USA:
ISO 9001 Certified
One Omega Drive, Box 4047
Stamford, CT 06907-0047
Tel: (203) 359-1660
e-mail: info@omega.com

FAX: (203) 359-7700

Canada:
976 Bergar
Laval (Quebec) H7L 5A1
Tel: (514) 856-6928
e-mail: canada@omega.com

FAX: (514) 856-6886

For immediate technical or application assistance:

USA and Canada: Sales Service: 1-800-826-6342 / 1-800-TC-OMEGA™
Customer Service: 1-800-622-2378 / 1-800-622-BEST™
Engineering Service: 1-800-872-9436 / 1-800-USA-WHEN™
TELEX: 996404 EASYLINK: 62968934 CABLE: OMEGA

Mexico and Latin America: Tel: (95) 800-TC-OMEGA™ FAX: (95) 203-359-7807
En Español: (203) 359-1660 ext: 2203 e-mail: espanol@omega.com

Servicing Europe:

Benelux: Postbus 8034, 1180 LA Amstelveen, The Netherlands
Tel: (31) 20 6418405 FAX: (31) 20 6434643
Toll Free in Benelux: 06 0993344
e-mail: nl@omega.com

Czech Republic: Ostravska 767, 733 01 Karvina
Tel: 420 (69) 6311627 FAX: 420 (69) 6311114
e-mail: czech@omega.com

France: 9, rue Denis Papin, 78190 Trappes
Tel: 33 0130-621-400 FAX: 33 0130-699-120
Toll Free in France: 0800-4-06342
e-mail: france@omega.com

Germany/Austria: Daimlerstrasse 26, D-75392 Deckenpfronn, Germany
Tel: 49 (07056) 3017 FAX: 49 (07056) 8540
Toll Free in Germany: 0130 11 21 66
e-mail: germany@omega.com

United Kingdom: 25 Swannington Road, P.O. Box 7, Omega Drive,
ISO 9002 Certified Broughton Astley, Leicestershire, Irlam, Manchester,
LE9 6TU, England M44 5EX, England
Tel: 44 (1455) 285520 Tel: 44 (161) 777-6611
FAX: 44 (1455) 283912 FAX: 44 (161) 777-6622
Toll Free in England: 0800-488-488
e-mail: uk@omega.com

It is the policy of OMEGA to comply with all worldwide safety and EMC/EMI regulations that apply. OMEGA is constantly pursuing certification of its products to the European New Approach Directives. OMEGA will add the CE mark to every appropriate device upon certification.

The information contained in this document is believed to be correct but OMEGA Engineering, Inc. accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.

WARNING: These products are not designed for use in, and should not be used for, patient connected applications.



WARRANTY/DISCLAIMER

OMEGA ENGINEERING, INC. warrants this unit to be free of defects in materials and workmanship for a period of **13 months** from date of purchase. OMEGA Warranty adds an additional one (1) month grace period to the normal **one (1) year product warranty** to cover handling and shipping time. This ensures that OMEGA's customers receive maximum coverage on each product.

If the unit should malfunction, it must be returned to the factory for evaluation. OMEGA's Customer Service Department will issue an Authorized Return (AR) number immediately upon phone or written request. Upon examination by OMEGA, if the unit is found to be defective it will be repaired or replaced at no charge. OMEGA's WARRANTY does not apply to defects resulting from any action of the purchaser, including but not limited to mishandling, improper interfacing, operation outside of design limits, improper repair, or unauthorized modification. This WARRANTY is VOID if the unit shows evidence of having been tampered with or shows evidence of being damaged as a result of excessive corrosion; or current, heat, moisture or vibration; improper specification; misapplication; misuse or other operating conditions outside of OMEGA's control. Components which wear are not warranted, including but not limited to contact points, fuses, and triacs.

OMEGA is pleased to offer suggestions on the use of its various products. However, OMEGA neither assumes responsibility for any omissions or errors nor assumes liability for any damages that result from the use of its products in accordance with information provided by OMEGA, either verbal or written. OMEGA warrants only that the parts manufactured by it will be as specified and free of defects. OMEGA MAKES NO OTHER WARRANTIES OR REPRESENTATIONS OF ANY KIND WHATSOEVER, EXPRESSED OR IMPLIED, EXCEPT THAT OF TITLE, AND ALL IMPLIED WARRANTIES INCLUDING ANY WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE HEREBY DISCLAIMED. LIMITATION OF LIABILITY: The remedies of purchaser set forth herein are exclusive and the total liability of OMEGA with respect to this order, whether based on contract, warranty, negligence, indemnification, strict liability or otherwise, shall not exceed the purchase price of the component upon which liability is based. In no event shall OMEGA be liable for consequential, incidental or special damages.

CONDITIONS: Equipment sold by OMEGA is not intended to be used, nor shall it be used: (1) as a "Basic Component" under 10 CFR 21 (NRC), used in or with any nuclear installation or activity; or (2) in medical applications or used on humans. Should any Product(s) be used in or with any nuclear installation or activity, medical application, used on humans, or misused in any way, OMEGA assumes no responsibility as set forth in our basic WARRANTY/DISCLAIMER language, and additionally, purchaser will indemnify OMEGA and hold OMEGA harmless from any liability or damage whatsoever arising out of the use of the Product(s) in such a manner.

RETURN REQUESTS / INQUIRIES

Direct all warranty and repair requests/inquiries to the OMEGA Customer Service Department. BEFORE RETURNING ANY PRODUCT(S) TO OMEGA, PURCHASER MUST OBTAIN AN AUTHORIZED RETURN (AR) NUMBER FROM OMEGA'S CUSTOMER SERVICE DEPARTMENT (IN ORDER TO AVOID PROCESSING DELAYS). The assigned AR number should then be marked on the outside of the return package and on any correspondence.

The purchaser is responsible for shipping charges, freight, insurance and proper packaging to prevent breakage in transit.

- FOR **WARRANTY** RETURNS, please have the following information available BEFORE contacting OMEGA:
1. P.O. number under which the product was PURCHASED,
 2. Model and serial number of the product under warranty, and
 3. Repair instructions and/or specific problems relative to the product.

- FOR **NON-WARRANTY** REPAIRS, consult OMEGA for current repair charges. Have the following information available BEFORE contacting OMEGA:
1. P.O. number to cover the COST of the repair,
 2. Model and serial number of product, and
 3. Repair instructions and/or specific problems relative to the product.

OMEGA's policy is to make running changes, not model changes, whenever an improvement is possible. This affords our customers the latest in technology and engineering.

OMEGA is a registered trademark of OMEGA ENGINEERING, INC.

© Copyright 1996 OMEGA ENGINEERING, INC. All rights reserved. This document may not be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form, in whole or in part, without prior written consent of OMEGA ENGINEERING, INC.