


HH804W

HH804WE WIRELESS DUAL INPUT RTD DIGITAL THERMOMETER

DIGITAL THERMOMETER INSTRUCTION M4552/0711

SHEET Shop online at: omega.com e-mail: info@omega.com For latest product manuals: omegamanual.info MADE IN TAIWAN

•1 YEAR





OMEGAnet[®] On-Line Service Internet e-mail omega.com info@omega.com

Servicing North America:

U.S.A: One Omega Drive, Box 4047 Stamford, CT 06907-0047 Tel: (203) 359-1660 FAX: (203) 359-7700 Certified e-mail: info@omega.com

FC

CE

Canada: 976 Bergar Laval (Quebec) H7L 5A1, Canada Tel: (514) 856-6928 FAX: (514) 856-6886 e-mail: info@omega.ca

For immediate technical or application assistance:

U.S.A Sales Service: 1-800-826-6342/1-800-TC-OMEGA®

 and
 Customer Service: 1-800-622-2378/1-800-622-BEST®

 Canada:
 Engineering Service: 1-800-872-9436/1-800-USA-WHEN®

 TELEX: 996404 EASYLINK: 62968934 CABLE: OMEGA

Mexico: En Espan ol: (001) 203-359-7803 e-mail: espanol@omega.com FAX: (001) 203-359-7807 info@omega.com.mx

Servicing Europe:

- Benelux:
 Postbus 8034, 1180 LA Amstelveen, The Netherlands Tel: +31 (0)20 3472121
 FAX: +31 (0)20 6434643 Toll Free in Benelux: 0800 0993344

 e-mail:
 sales@omegaeng.nl
- Czech
 Frystatska 184, 733 01 Karviná, Czech Republic

 Republic:
 Tel: +420 (0)59 6311899
 FAX: +420 (0)59 6311114

 Toll Free: 0800-1-66342
 e-mail: info@omegashop.cz
- France: 11, rue Jacques Cartier, 78280 Guyancourt, France Tel: +33 (0)1 61 37 2900 FAX: +33 (0)1 30 57 5427 Toll Free in France: 0800 466 342 e-mail: sales@omega.fr
- Germany/ Daimlerstrasse 26, D-75392 Deckenpfronn, Germany Austria: Tel: + 49 (0)7056 9398-0 FAX: +49 (0)7056 9398-29 Toll Free in Germany: 0800 639 7678 e-mail: info@omega.de
- United One Omega Drive, River Bend Technology Centre
- Kingdom: Northbank, Irlam, Manchester
- ISO 9002 M44 5BD United Kingdom
- <u>Certified</u> Tel: +44 (0)161 777 6611 FAX: +44 (0)161 777 6622 Toll Free in United Kingdom: 0800-488-488 e-mail: sales@omega.co.uk

It is the policy of OMEGA Engineering, Inc. to comply with all worldwide safety and EMC/EMI regulations that apply. OMEGA is constantly pursuing certification of its products to the European New Approach Directives. OMEGA will add the CE mark to every appropriate device upon certification. The information contained in this document is believed to be correct, but OMEGA accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.

WARNING: These products are not designed for use in, and should not be used for, human applications.

INTRODUCTION

This instrument is a $4\frac{1}{2}$ digit, compact-sized portable digital thermometer designed to use external 100 Ω Platinum RTD as temperature sensor. Temperature indication follows Reference Temperature/Resistance Tables (Pt385 for European Curve, Alpha=.00385. Pt3926 for American Curve, Alpha=.003926. Pt3916 for Japan Curve, Alpha=.003916.)

SAFETY INFORMATION

It is recommended that you read the safety and operation instructions before using the thermometer.

WARNING

To avoid electrical shock, do not use this instrument when working voltages at the measurement surface over 24V AC or DC.

WARNING To avoid damage or burns, do not make temperature measurement in microwave ovens.

FEDERAL COMMUNICATIONS COMMISSION

This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:(1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

NOTE

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection. This equipment generates, uses and can radiated radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit dif-
- ferent from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Shielded interface cables must be used in order to comply with emission limits.

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment.

WIRELESS NOTE

Wireless receiver must keep a distance at least 40cm from the meter and meter to meter distance must be at least 30cm.

SPECIFICATIONS

ELECTRICAL

Temperature Scale: Celsius or Fahrenheit user-selectable Measurement Range:

Pt385(100Ω) -200°C to 800°C, (-328°F to 1472°F) Pt3916/Pt3926(100Ω) -200°C to 630°C, (-328°F to 1166°F)

Resolution: 0.1°C or 0.2°F

Accuracy:

Accuracy is specified for operating temperatures over the range of 18° C to 28° C (64° F to 82° F), for 1 year, not including RTD probe error.

\pm (0.05% rdg + 0.2°C) on °C scale

 $\pm (0.05\% \text{ rdg} + 0.4^{\circ}\text{F}) \text{ on }^{\circ}\text{F} \text{ scale}$

Temperature Coefficient:

0.1 times the applicable accuracy specification per $^{\circ}$ C from 0 $^{\circ}$ C to 18 $^{\circ}$ C and 28 $^{\circ}$ C to 50 $^{\circ}$ C (32 $^{\circ}$ F to 64 $^{\circ}$ F and 82 $^{\circ}$ F to 122 $^{\circ}$ F).

Input Protection:

24V dc or 24V ac rms maximum input voltage on any combination of input pins.

Maximum Differential Common Mode Voltage (Maximum Voltage between T1 and T2 during measurement): 1volt.

Reading Rate: 1 time per second.

Input Connector:

Accepts for RTD 3 wires or 4 wires subminiature 4-Prong type connectors (flat blades spaced 7.9mm, center to center).

(9) Wireless Features:

Frequency range: 910~920MHz (HH804W) 868.1~868.5MHz (HH804WE)

Low current consumption less than 1mA. The transmitting distance can reach 25M without magnetic interference.

ENVIRONMENTAL

Ambient Operating Ranges:

0°C to 50°C (32°F to 122°F) <80% R.H. **Storage Temperature:** -20°C to 60°C (-4°F to 140°F) <70% R.H.

GENERAL

Display:

4½ digit liquid crystal display (LCD) with maximum reading of 19999. **Overload:** "-----", or "OL" is display. **Battery:** 1.5V x 4 PCS (SIZE AAA) UM-4 R03. **Battery Life:**

200 hours typical with carbon zinc battery. **Auto power off:** 30 minutes, press power key to resume operation.

Dimensions:

160mm(H) x83mm(W) x 38mm(D). **Weight:** Approx. 255g including batteries.

OPERATING INSTRUCTIONS

1. "⁽⁾" Power Button

The "①" key turns the thermometer on or off. In the SET mode cannot be powered off. Exit SET mode to power off.

APO function mode

Press "①" power key for more than 6 seconds to disable the auto-power function. The display will show "APO OFF".

2. °C/°F Selecting the Temperature Scale (Main display)

Reading the main displayed in either degrees $Celsius(^{\circ}C)$ or degrees Fahrenheit($^{\circ}F$). When the thermometer is turned on, it is set to the temperature scale that was in use when the thermometer was last turned off. To change the temperature scale, press the " $^{\circ}C/^{\circ}F$ " key.

3. "Q" Display Back-Light

Press the "Q" key to turn on or turn off the Back-Light.

4. T1 T2/T1-T2 Main display Input Selection

The input selection indicates which input is selected for main display; T1 probe, T2 probe or the difference between the two probes (T1-T2). When the thermometer is turned on, it is set to T1, when main display input selected T1, then T1 input can select alternate of probe by pressing Pt385/Pt3926/Pt3916 key switch.

5. TYPE(Pt385/Pt3926/Pt3916) Input RTD Probe Select (only Main display)

The "TYPE" key switches the T1 input to select the Pt385, Pt3926 or Pt3916 RTD probe as input, when main display input selected T1. When the thermometer is turned on, it is set to the probe selected that was in use when the thermometer was last turned off.

6. MIN/MAX with Time record mode (only Main display)

Press "MIN/MAX" key to enter the MIN MAX Recording mode, (displays the Maximum reading with time, Minimum reading with time and Average reading stored in record mode). In the this mode the automatic power-off feature is disabled and "D" key, "°C/°F" key, "REL" key, "SET" key, "Hi/Lo Limits" key and main display "T1 T2 T1-T2" key, "TYPE" key are disabled. The beeper emits a tone when a new minimum or maximum value is recorded.

Press "MIN/MAX" key to cycle through the MAX, MIN and AVG readings. If an overload is recorded, the averaging function is stopped. In this mode, press the "HOLD" key to stop the recording of readings, all values are held, press again to restart recording.

To prevent accidental loss of MIN, MAX and AVG data, this mode can only be cancelled by pressing and holding the "MIN/MAX" key for 2 seconds. All recorded readings are erased.

7. REL Relative mode (only Main display)

Press the "REL" key to enter the relative mode, zero the display, and store the displayed Reading as a reference value. REL is shown on the display. Press "REL" key again to exit the relative mode. The relative reference value can also be entered by the user. (See "SET mode" later in this manual). When the desired relative value has been entered, press "REL" key to enter the relative mode and than press "SET" key use the entered relative value as a reference value. Press "REL" key again to exit the relative mode. In the relative mode, the value (can not >±1999.9 counts) shown on the LCD is always the difference between the stored reference and the present reading.

8. HOLD mode (only Main display)

Press the "HOLD" key to enter the Data Hold mode, the "HOLD" annunciator is displayed. When HOLD mode is selected, the thermometer held the present readings and stops all further measurements. Press the "HOLD" key again to cancel HOLD mode causing thermometer to resume taking measurements. In the MIN/MAX recording mode, press "HOLD" key to stop the recording. Press "HOLD" key again to resume recording. (Previously recorded read are not erased).

9. SET mode (Relative value set, Time set and Hi/Lo Limits value set)

- **9.1** Press the "SET" key to enter relative values SET mode (Press "ENTER" key to skip setting relative value). = = = =.= is displayed on the main display. The relative value is entered via overlay numbers, press overlay "ENTER" key to store the relative value, and advance to elapsed time set mode.
- **9.2** Elapsed time set mode, (Press ENTER key to skip Elapsed time set mode) =.= = = : = = is displayed in second and third display. Time (hours, minutes, seconds) value is entered via overlay numbers, press overlay "ENTER" key to store time value. Elapsed time starts from set time value.
- 9.3 Hi Limit value set mode, (Press "ENTER" key to skip Hi Limit value set mode), = = = =.= is displayed in main display, Hi Limit value is entered via overlay numbers, then press "ENTER" key to store the Hi Limit value. = = = =.= is displayed in main display, Lo Limit value is entered via overlay numbers, then press overlay "ENTER" key to store the Lo Limit value and exit SET mode.
- **9.4** When the thermometer is turned on it uses the relative value and Hi/Lo Limits values that were entered when thermometer was last in use.

10. T1/T2 T1-T2 second display Input Selection

The input selection indicates which input is selected for second display; T1 probe, T2 probe or the difference between the two probes (T1-T2), when the thermometer is turned on, it is set to T2 input can select alternate of probe by second display Pt385/Pt3926/Pt3916 key switch.

11. TYPE(Pt385/Pt3926/Pt3916) Input RTD Probe select (second display)

The Pt385/Pt3926 key switches the T2 input to select the Pt385, Pt3926 or Pt3916 RTD probe as input, when second display input selected T2. When the thermometer is turned on, it is set to the RTD selected that was in use when the thermometer was last turned off.

12. Hi/Lo Limits mode (only Main display)

Press "Hi/Lo Limits" key to enter the Hi/Lo Limits comparative mode. When the input temperature value exceeds the Hi or Lo Limits value, the beeper emits a continuous or pulsed tone. Press "Hi/Lo Limits" key again to exit the Hi/Lo Limits mode.

13. WIRELESS MODE:

Press the " $\binom{(n)}{k}$ " key for more than two seconds to start wireless function. Press the " $\binom{(n)}{k}$ " key again for another two seconds to stop wireless function. The wireless mode will shut down if there is no wireless signal for two minutes.

To SET CH/ID to 00,00, press the "Hi/Lo Limits" key and "①" power key for more than 6 seconds with the meter powered down. The meter will set channel and ID to 00,00 status. The second display will show 00, which means that the channel and ID has been set to 00.

To check the channel and ID of the meter:

When the meter is off, press "°C/°F" key and " \bigcirc " for 5 seconds, LCD's main display will show channel number, the second display will show ID number.

RTD PROBE CONNECTION

P/N: 14061-0800-9 MINI DIN PLUG Rear view Front view - (1) **-** (1) + (1) 3 Wires 4 Wires [G] [+1 + (1) + (2) + (2) $\overline{\mathbf{A}}$ RTD RTD 100Ω 100 Q (Pt385/Pt3916/Pt3926) (Pt385/Pt3916/Pt3926) - (2) [-] - (1) - (2)

TEMPERATURE VSRESISTANCE TABLE (ITS90)

°C	Pt385	Pt3926	Pt3916
-200°C	18.521Ω	16.996Ω	17.057Ω
-100°C	60.256Ω	59.479Ω	59.565Ω
0°C	100.000Ω	100.000Ω	100.000Ω
100°C	138.505Ω	139.272Ω	139.171Ω
200°C	175.856Ω	177.362Ω	177.155Ω
300°C	212.052Ω	214.275Ω	213.957Ω
400°C	247.092Ω	250.018Ω	249.584Ω
500°C	280.977Ω	284.591Ω	284.036Ω
600°C	313.708Ω	317.994Ω	317.313Ω
700°C	345.280Ω	-	-
800°C	375.700Ω	-	-

OPERATOR MAINTENANCE

WARNING

To avoid possible electrical shock, disconnect the thermocouple connectors from the thermometer before removing the cover.

Battery Replacement

- 1. Power is supplied by 4pcs 1.5V (AAA SIZE) UM-4 R03.
- 2. The "-" appears on the LCD display when replacement is needed. To replace battery remove screw from back of meter and lift off the battery cover.
- 3. Remove the battery from battery contacts and replace.
- 4. When not in use for long periods remove battery.
- 5. Do not store in locations with high temperatures, or high humidity.

Cleaning

Periodically wipe the case with a damp cloth and detergent, do not use abrasives or solvents.

*Software Operation manual is on the Software disk.

WARRANTY/DISCLAIMER

OMEGA ENGINEERING, INC, warrants this unit to be free of defects in materials and workmanship for a period of **13 months** from date of purchase. OMEGA's WARRANTY adds an additional one (1) month grace period to the normal **one (1) year product warranty** to cover handling and shipping time. This ensures that OMEGA's customers receive maximum coverage on each product.

If the unit malfunctions, it must be returned to the factory for evaluation. OMEGA's Customer Service Department will issue an Authorized Return (AR) number immediately upon phone or written request. Upon examination by Miniper infineurately open profile of writer i regless. Open examination by OMEGA, if the unit is found to be defective, it will be repaired or replaced at no charge. OMEGA's WARRANTY does not apply to defects resulting from any action of the purchaser, including but not limited to mishandling, improper interfacing, operation outside of design limits, improper repair, or unauthorized modification. This WARRANTY is VOID if the unit shows evidence of having been tampered with or shows evidence of having been damaged as a result of excessive corrosion; or current, heat, moisture or vibration; improper specification; misapplication; misuse or other operating conditions outside of OMEGA's control. Components in which wear is not warranted, include but are not limited to contact points, fuses, and triacs.

OMEGA is pleased to offer suggestions on the use of its various products. However, OMEGA neither assumes responsibility for any omissions or errors nor assumes liability for any damages that result from the use of its products in accordance with information provided by OMEGA, either verbal or written. OMEGA warrants only that the parts manufactured by the company will be as Specified and free of defects. OMEGA MAKES NO OTHER WARRANTIES OR REPRESENTATIONS OF ANY KIND WHATSOEVER, EXPRESSED OR IMPLIED, EXCEPT THAT OF TITLE, AND ALL IMPLIED WARRANTIES INCLUDING ANY WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE HEREBY DISCLAIMED. LIMITATION OF LIABILITY: The remedies of purchaser set forth herein are exclusive, and the total liability of OMEGA with respect to this order, whether based on contract, warranty, negligence, indemnification, strict liability or otherwise, shall not exceed the purchase price of the component upon which liability is based. In no event shall OMEGA be liable for consequential, incidental or special damages.

CONDITIONS: Equipment sold by OMEGA is not intended to be used, nor shall it be used: (1) as a "Basic Component" under 10 CFR 21 (NRC), used in or with any nuclear installation or activity; or (2) in medical applications or used on humans. Should any Product(s) be used in or with any nuclear installation or activity, medical application, used on humans, or misused in any way, OMEGA assumes no responsibility as set forth in our basic WARRANTY/DISCLAIMER language, and, additionally, purchaser will indemnify OMEGA and hold OMEGA harmless from any liability or damage whatsoever arising out of the use of the Product(s) in such a manner.

RETURN REQUESTS / INQUIRIES

Direct all warranty and repair requests/inquiries to the OMEGA Customer Service Department. BEFORE RETURNING ANY PRODUCT(S) TO OMEGA, PURCHASER MUST OBTAIN AN AUTHORIZED RETURN (AR) NUMBER FROM OMEGA'S CUSTOMER SERVICE DEPARTMENT (IN ORDER TO AVOID PROCESSING DELAYS). The assigned AR number should then be marked on the outside of the return package and on any correspondence. The purchaser is responsible for shipping charges, freight, insurance and proper packaging to prevent breakage in transit.

FOR WARRANTY RETURNS, please	FOR NON-WARRANTY REPAIRS,
have the following information	consult OMEGA for current repair
available BEFORE contacting	charges. Have the following
OMEGA:	information available BEFORE
1. Purchase Order number under	contacting OMEGA:
which the product was PURCHASED,	 Purchase Order number to cover the COST of the repair,
2. Model and serial number of the product under warranty, and	Model and serial number of the product, and
2 Papair instructions and/or apositio	2 Papair instructions and/or aposifia

3. Repair instructions and/or specific 3. Repair instructions and/or specifi problems relative to the product. problems relative to the product.

OMEGA's policy is to make running changes, not model changes, wheneve an improvement is possible. This affords our customers the latest in technology and engineering.

OMEGA is a registered trademark of OMEGA ENGINEERING, INC. © Copyright 2011 OMEGA ENGINEERING, INC. All rights reserved. This document may not be copied, photocopied, reproduced, translated, or reduced to any electronic medium or machine-readable form, in whole or in part, without the prior written consent of OMEGA ENGINEERING, INC.

Where Do I Find Everything I Need for **Process Measurement and Control?** OMEGA...Of Course! Shop online at omega.com®

TEMPERATURE

- Thermocouple, RTD & Thermistor Probes,
- Connectors, Panels & Assemblies Wire: Thermocouple, RTD & Thermistor
- Calibrators & Ice Point References
- Recorders, Controllers & Process Monitors
- Infrared Pyrometers

PRESSURE, STRAIN AND FORCE

- ☑ Transducers & Strain Gages
- Load Cells & Pressure Gages
- Displacement Transducers
- Instrumentation & Accessories

FLOW/LEVEL

- ☑ Rotameters, Gas Mass Flowmeters & Flow Computers
- Air Velocity Indicators
- Turbine/Paddlewheel Systems ☑ Totalizers & Batch Controllers

pH/CONDUCTIVITY

- ☑ pH Electrodes, Testers & Accessories
- Benchtop/Laboratory Meters
- Controllers, Calibrators, Simulators & Pumps
- Industrial pH & Conductivity Equipment

DATA ACQUISITION

- Data Acquisition & Engineering Software
- Communications-Based Acquisition Systems
- Plug-in Cards for Apple, IBM & Compatibles
- Data Logging Systems
 Recorders, Printers & Plotters

HEATERS

- ☑ Heating Cable
- Cartridge & Strip Heaters
- Immersion & Band Heaters
- Flexible Heaters
- Laboratory Heaters

ENVIRONMENTAL MONITORING AND CONTROL

- Metering & Control Instrumentation
- Refractometers
- Pumps & Tubing
- Air, Soil & Water Monitors
- Industrial Water & Wastewater Treatment
- PH, Conductivity & Dissolved Oxygen Instruments